

Sarah Jones
(660) 785-4159
FAX: (660) 785-7337



McClain Hall 106
100 E Normal
Kirksville, MO 63501-4221

TO: ALL PROSPECTIVE VENDORS

RE: Addendum No. 1
Glove Box
Project No. 2026-04

Addendum No. 1 is issued to inform vendors that we have received questions that require a response, and therefore, an addendum. Please refer to the questions and responses provided.

Please confirm that we have the correct email address for your company. This addendum and subsequent addendums will be sent to all firms who have expressed their interest in submitting a proposal.

Vendors are to acknowledge receipt of Addendum No. 1 in their response.

If you have further questions, please contact me at sjones@truman.edu. Thank you.

Sincerely,

A handwritten signature in cursive script that reads "Sarah Jones".

Sarah Jones
Purchasing Buyer

Project No. 2026-04 Glove Box Addendum 1

1. Are we able to know what materials will be used within the glove box to ensure any substitutes in glove material availability or lighting source (LED vs fluorescence) will not have a negative impact on the use of the equipment?
 - a. Mostly, the materials used within the glove box will be air-sensitive or moisture-sensitive substances like organometallic perovskites, battery materials, etc.
2. Can butyl be substituted with another material such as neoprene??
 - a. Truman will accept neoprene as a substitute for butyl gloves.
3. With respect to the size requirement, is the space restricted to accommodate only a compact unit, or would something with larger dimensions be acceptable?
 - a. Truman will prefer a compact and portable unit.
4. Are there any requirements for this to be a standalone unit vs one that requires a base/stand?
 - a. There are no such requirements.